



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: LT45
Description: Light Meter
Date of Issue: 21-Apr-15

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC
Report Number: R09020304E-A02
Report Date of Issue: 6/8/2012

Standards:

EN 61326-1:2006
CISPR11 Amend.2:2006 Group 1 Class B
EN 61326-1:2006
IEC 61000-4-2:2008
IEC 61000-4-3:2006+A1:2007

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive and RoHS Directive.



Mark Sultzbach / QA Administrator
(for Tony Campagna / Director of Quality Services)